

## **IECEx Certificate** of Conformity

## INTERNATIONAL ELECTROTECHNICAL COMMISSION **IEC Certification Scheme for Explosive Atmospheres**

for rules and details of the IECEx Scheme visit www.iecex.com

Cert	ifi	ca	te	N	0	

**IECEx BVS 11.0088X** 

issue No.:1

Certificate history:

Status

Current

Issue No. 1 (2013-7-1) Issue No. 0 (2011-11-25)

Date of Issue:

2013-07-01

Page 1 of 4

Applicant:

Heinrichs Messtechnik GmbH

Robert-Perthel-Strasse 9

50739 Köln Germany

Electrical Apparatus:

Optional accessory:

Mass flow meter converter Type UMC4-\*\*\*\*\*

Type of Protection:

Flameproof enclosure 'd', Equipment protection by intrinsic safety "i", Equipment with

equipment protection level (EPL) Ga

Marking:

Ex d [ia Ga] IIC T4-T3 Gb

Approved for issue on behalf of the IECEx Certification Body:

H.-Ch. Simanski

Position:

Head of Certification Body

Signature:

(for printed version)

Date:

1. This certificate and schedule may only be reproduced in full.

2. This certificate is not transferable and remains the property of the issuing body.

3. The Status and authenticity of this certificate may be verified by visiting the Official IECEx Website.

Certificate issued by:

**DEKRA EXAM GmbH** Dinnendahlstrasse 9 44809 Bochum Germany





## **IECEx Certificate** of Conformity

Certificate No.:

IECEx BVS 11.0088X

Date of Issue:

2013-07-01

Issue No.: 1

Page 2 of 4

Manufacturer:

Heinrichs Messtechnik GmbH

Robert-Perthel-Strasse 9

50739 Köln Germany

Additional Manufacturing location

This certificate is issued as verification that a sample(s), representative of production, was assessed and tested and found to comply with the IEC Standard list below and that the manufacturer's quality system, relating to the Ex products covered by this certificate, was assessed and found to comply with the IECEx Quality system requirements. This certificate is granted subject to the conditions as set out in IECEx Scheme Rules, IECEx 02 and Operational Documents as amended.

#### STANDARDS:

The electrical apparatus and any acceptable variations to it specified in the schedule of this certificate and the identified documents, was found to comply with the following standards:

IEC 60079-0: 2011

Explosive atmospheres - Part 0: General requirements

Edition: 6.0

IEC 60079-1: 2007-04

Explosive atmospheres - Part 1: Equipment protection by flameproof enclosures "d"

Edition: 6

IEC 60079-11: 2011

Explosive atmospheres - Part 11: Equipment protection by intrinsic safety "i"

Edition: 6.0 IEC 60079-26: 2006

Edition: 2

Explosive atmospheres - Part 26: Equipment with equipment protection level (EPL) Ga

This Certificate does not indicate compliance with electrical safety and performance requirements other than those expressly included in the Standards listed above.

## **TEST & ASSESSMENT REPORTS:**

A sample(s) of the equipment listed has successfully met the examination and test requirements as recorded in

Test Report: DE/BVS/ExTR11.0117/01

**Quality Assessment Report:** 

DE/BVS/QAR11.0001/01



# IECEx Certificate of Conformity

Certificate No.:

IECEx BVS 11.0088X

Date of Issue:

2013-07-01

Issue No.: 1

Page 3 of 4

#### Schedule

#### EQUIPMENT:

Equipment and systems covered by this certificate are as follows:

#### Description

The mass flow meter converter is used in combination with a mass flow sensor for measurement of mass flow of liquids and gases in pipes.

The converter consists of an enclosure type XD-ID100win (IECEx FTZU 10.0019U) and the electronic devices fixed inside.

The signal circuits can be executed either as intrinsically safe Ex ia (terminals 11 - 20) in type UMC4-\*\*\*\*1 or as non-intrinsically safe (terminals 41 - 50) in type UMC4-\*\*\*\*2.

The Exciter circuit (terminals 9 and 10), the temperature sensor circuit (terminals 5 up to 8) and the sensor circuits (terminals 1 - 2 and 3 - 4) are always intrinsically safe, level of protection Ex ia.

#### Type designation

See Annex

#### **Parameters**

See Annex

#### CONDITIONS OF CERTIFICATION: YES as shown below:

- 1 As cable entries separately certified flameproof cable glands have to be used.
- 2 If the mass flow meter converter is connected by conduit entries they have to be certified for this purpose and the associated stopping boxes have to be mounted immediately to the enclosure.
- 3 The correlation between ambient temperature range, process temperature and temperature class is shown in the manufacturer's instructions.
- 4 If sensor and mass flow meter are installed separately, it must be ensured that between sensor and flow meter potential equalisation is arranged.
- 5 For type UMC4-E\*\*\*\* the cable to the sensor has to be installed in a way that tensile force is omitted.



# IECEx Certificate of Conformity

Certificate No.:

IECEx BVS 11.0088X

Date of Issue:

2013-07-01

Issue No.: 1

Page 4 of 4

## DETAILS OF CERTIFICATE CHANGES (for issues 1 and above):

The mass flow meter converter has been assessed in acc. with actual standard versions.

Annexe: BVS\_11\_0088x\_Heinrichs\_Annex.pdf



# IECEx Certificate DEKRA of Conformity



**Certificate No.:** 

IECEx BVS 11.0088X Annex

Page 1 of 3

## Type designation

Mass flow meter converter type UMC4-\*\*\*\*\* Type UMC4-\*1\*A\*\*

0 = without

Type of protection signal circuits

0 = without

1 = intrinsically safe Ex ia 2 = non-intrinsically safe

Certificate

2 = Ex d [ia Ga] IIC T4-T3 Gb

A = Output circuits

Power supply

1 = 90 - 265 V AC

2 = 19 - 36 V DC, 24 V AC

1 = inclusive control unit

Mounting option

B = Compact version

D = Remote version with connection box

E = Remote version with permanently connected cable

## **Parameters**

Power circuit (terminals L, N and PE)

Type UMC4-**1*** Nominal voltage Max. voltage	Um	AC AC	90 - 265 265	V V
Type UMC4-**2***			•	
Nominal voltage		AC	24	V
Max. voltage	Um	AC	60	V
Nominal voltage		DC	19 - 36	V
Max. voltage	Um	DC	60	V

2 Passive non intrinsically safe circuits

Type UMC4-\*\*\*\*2

Current output 1 (terminals 41 - 42) Current output 2 (terminals 43 - 44) Impulse output (terminals 46 - 47) Status output (terminals 49 - 50)

Max. voltage	Um	AC/DC	60	V
Max. current of the power supply			500	mA



# IECEx Certificate DEKRA of Conformity



Certificate No.: IECEx BVS 11.0088X Annex

Page 2 of 3

3 3.1	Sensor circuits type of protection Ex ia II Exciter circuit (terminals 9 and 10) Voltage Current Power Linear output characteristic	Uo Io Po	DC	12.15 90 271	V mA mW
	Type of protection Ex ia IIC Max. external inductance Max. external capacitance	Lo Co		5 1320	mH nF
	Type of protection Ex ia IIB Max. external inductance Max. external capacitance	Lo Co		18 8400	mH nF
3.2	Temperature sensor circuit (terminals 5 up to 8) Voltage Current Power Linear output characteristic	Uo Io Po	DC	12.15 3.84 12	V mA mW
	Type of protection Ex ia IIC Max. external inductance Max. external capacitance	Lo Co		1000 1305	mH nF
	Type of protection Ex ia IIB Max. external inductance Max. external capacitance	Lo Co		1000 8385	mH nF
3.3	Sensor circuit (terminals 1 - 2 and 3 - 4) Values for each circuit Voltage Current Power Linear output characteristic	Uo Io Po	DC	12.15 16 47	V mA mW
	Type of protection Ex ia IIC Max. external inductance Max. external capacitance	Lo Co		140 1305	mH nF
	Type of protection Ex ia IIB Max. external inductance Max. external capacitance	Lo Co		510 8385	mH nF
	For type UMC4-E**** the following values for the permanent Cable capacitance Cable inductance	connec	cted cab	100	oe regarded: pF/m µH/m



# IECEx Certificate DEKRA of Conformity



**Certificate No.:** 

**IECEx BVS 11.0088X Annex** 

Page 3 of 3

3.4 Current output 1 (terminals 11 - 12) and

Current output 2 (terminals 13 - 14)

Type of protection Ex ia II

Passive circuit

Values for each circuit

Voltage	Ui	DC	30	V
Current	li		150	mΑ
Power	Pi		1.3	W
Effective internal inductance	Li		0.1	mΗ
Effective internal capacitance	Ci		20	nF

3.5 Impulse output (terminals 16 - 17) and

Status output 2 (terminals 19 -20)

Floating optocoupler output circuit type of protection Ex ia IIC

Voltage	Ui	DC	30	V
Current	li	2	00	mA
Power	Pi		3	W
Effective internal inductance	Li	negligible		
Effective internal capacitance	Ci	negligible		

3.6 Ambient temperature range Ta

Depending on the process temperature and the temperature class shown in the following table:

Process temperature -20 °C up to	Ambient temperature -20 °C up to	Temperature class
100 °C	60 °C	T4
130 °C	55 °C	T4
150 °C	50 °C	T3

If the converter is mounted remote from the process the ambient temperature range is -20 °C up to +60 °C.